

Oxygenation adjustment of ventilator combined with venovenous extracorporeal membrane oxygenation

Chuanzhu Lv^a, Wei Guo^{b,*}, Xufeng Chen^c, Yong Mei^c, Jing Sun^d, Ling Wang^e, Yuguo Chen^f, Huadong Zhu^g, Yu Cao^h, Yanfan Chaiⁱ, and Emergency Medicine Branch of Chinese Geriatrics Society

Abstract

In recent years, venovenous extracorporeal membrane oxygenation (VV-ECMO) has emerged as a critical intervention in the management of adult respiratory failure, with its clinical application expanding at a rapid pace annually. For critically ill patients receiving VV-ECMO support, achieving and maintaining a balance between oxygen supply and oxygen consumption is of paramount importance. Oxygen supply primarily relies on cardiac output and arterial oxygen content; notably, in patients under VV-ECMO support, arterial oxygen content is closely linked to the parameter configurations of both VV-ECMO and mechanical ventilators. Therefore, based on national and industrial needs, this document has been formulated by integrating the latest advancements and practical experience in this field from both domestic and international sources. Its core objectives are to standardize the adjustment of oxygen supply during the combined use of mechanical ventilators and VV-ECMO, provide unambiguous guidance for healthcare professionals, and ultimately enhance the success rate in the treatment of severe respiratory conditions.

Keywords: Adjustment, Mechanical ventilation, Oxygen supply, Respiratory failure, VV-ECMO

As the use of venovenous extracorporeal membrane oxygenation (VV-ECMO) for adult respiratory failure increases, managing oxygen delivery is critical. A total of 49 experts in the field of

emergency medicine drafted this guideline for regulating oxygen supply during combined VV-ECMO and mechanical ventilator support. Based on the latest evidence, it aims to provide clinicians with clear guidance to improve patient outcomes.

The datasets generated during and/or analyzed during the current study are available from the corresponding author on reasonable request.

^aDepartment of Emergency Medicine, Sichuan Academy of Medical Sciences & Sichuan Provincial Peoples' Hospital, Chengdu, Sichuan, China, ^bDepartment of Emergency Medicine, Beijing Hospital of Traditional Chinese Medicine, Capital Medical University, Beijing, China, ^cDepartment of Emergency Medicine, The First Affiliated Hospital of Nanjing Medical University, Nanjing, Jiangsu, China, ^dDepartment of Emergency Medicine, First Medical Center, Chinese PLA General Hospital, Beijing, China, ^eDepartment of Emergency Medicine, Shenzhen Traditional Chinese Medicine Hospital, Shenzhen, China, ^fDepartment of Emergency Medicine, Qilu Hospital of Shandong University, Jinan, Shandong, China, ^gDepartment of Emergency Medicine, Peking Union Medical College Hospital, Beijing, China, ^hDepartment of Emergency Medicine, West China Hospital, Sichuan University, Chengdu, Sichuan, China, ⁱDepartment of Emergency Medicine, Tianjin Medical University General Hospital, Tianjin, China.

*Corresponding author. Address: Department of Emergency Medicine, Beijing Hospital of Traditional Chinese Medicine, Capital Medical University, No.23, Meishuguan Back Street, Beijing 100010, China. E-mail address: guowei1010@126.com (W. Guo).

English version of this article is co-published in *Journal of Acute Disease* (doi:10.4103/jad.jad_123_25) and *Emergency and Critical Care Medicine* (doi:10.1097/EC9.000000000000174). Chinese version of this article is co-published in *Journal of Clinical Emergency* (doi:110.13201/j.jissn.1009-5918.2025.10.001) and *Chinese Journal of Critical Care Medicine* (doi:10.3969/j.issn.1002-1949.2025.10.002).

Copyright © 2026 The Author(s). Published by Wolters Kluwer Health, Inc.

This is an open access article distributed under the Creative Commons Attribution License 4.0 (CCBY), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Emergency and Critical Care Medicine (2026) 6:1

Received: 31 August 2025; Accepted: 27 November 2025.

Published online: 30 March 2026

<http://dx.doi.org/10.1097/EC9.000000000000174>

Scope

This document outlines the basic conditions, indications, contraindications, and oxygenation management protocol for the combined use of mechanical ventilation and VV-ECMO.

This document applies to oxygenation management by medical staff using ventilators in conjunction with VV-ECMO.

This document is not applicable to pediatric and infant patients.

Normative references

The document has no normative references.

Terms and definitions

The following terms and definitions apply to this document.

Venovenous extracorporeal membrane oxygenation^[1]

This technique involves draining venous blood from the body via percutaneous cannulation or surgical cutdown; the blood is then oxygenated and cleared of carbon dioxide by a membrane oxygenator before being infused back into the patient's venous system via a centrifugal pump, thereby assuming the gas exchange function and providing respiratory support.

Note: Refer to the “*Technical Operation Standard for Adult Extracorporeal Membrane Oxygenation*” (2024 Edition).^[2]

Positive end-expiratory pressure

Positive end-expiratory pressure (PEEP) is a technique used in mechanical ventilation. It maintains positive pressure in the

airways at the end of exhalation to prevent collapse of the small airways and alveoli and to improve functional residual capacity.

Source: WS 392-2012, 3.14, with revisions.^[3]

Predicted body weight

Predicted body weight (PBW) is a theoretical standard of healthy body weight used to assess nutritional status, guide drug dosage calculations, and set mechanical ventilation parameters.

Note 1: Formula: Adult male: $PBW = 50 + 0.91 \times [\text{height (cm)} - 152.4]$; Adult female: $PBW = 45.5 + 0.91 \times [\text{height (cm)} - 152.4]$.

Note 2: Refer to the “*Perioperative mechanical ventilation for adult non-cardiothoracic surgical patients: A Delphi consensus*” (2024 Edition).^[4]

Abbreviations

The following abbreviations apply to this document.

FiO₂: fraction of inspired oxygen
Hb: hemoglobin
PaCO₂: arterial carbon dioxide pressure
PaO₂: arterial oxygen pressure
pH: potential of hydrogen
SaO₂: arterial oxygen saturation
SpO₂: pulse oxygen saturation
SvO₂: venous oxygen saturation

Basic conditions^[5]

Environment, facilities, and equipment

A unit must have at least the following equipments: blood gas analyzer, hemodynamic monitoring devices, invasive ventilators, and continuous renal replacement therapy equipment, and be set up in an intensive care unit or equivalent resuscitation facilities.

All VV-ECMO devices, consumables, and related supporting facilities must be certified as compliant with technical and safety standards.

Personnel and team

Physicians performing VV-ECMO procedures must possess the following qualifications: (1) At least 5 years of clinical experience; (2) Professional qualifications in one of the following specialties: emergency medicine, critical care medicine, internal medicine, or surgery; and (3) Completion of specialized VV-ECMO training and the corresponding assessment.

Nurses performing VV-ECMO procedures must undergo systematic training and pass the assessment.

The VV-ECMO team shall consist of at least 3 members, including 2 licensed physicians and 1 nurse. Among the 2 licensed physicians, one should be the ECMO team leader, who is a physician with experience in VV-ECMO operation and holds a professional title of deputy chief physician or a higher rank. (1) Physicians are responsible for: indications assessment, catheterization operations, formulation and operation management of VV-ECMO treatment strategies (formulation of anticoagulation plans, hemodynamic regulation), weaning from the machine and complications management, etc. (2) Nurses are responsible for: equipment management, parameter monitoring, anticoagulation implementation, complications prevention, meticulous patient care, transfer and emergency response, etc.

Indications and contraindications^[6, 7]

Indications

Patients with respiratory failure who still have poor treatment effects under optimal mechanical ventilation conditions (FiO₂ ≥ 80%, tidal volume 4–6 mL/kg PBW, PEEP ≥ 10 cmH₂O) despite the combined use of protective lung ventilation strategies, lung recruitment maneuvers, prone position ventilation, etc, with a PaO₂/FiO₂ ratio < 80 mm Hg.

Patients with respiratory failure who still have hypercapnia, with a pH ≤ 7.25 and PaCO₂ ≥ 60 mm Hg, even after optimizing mechanical ventilation (respiratory rate ≥ 35 breaths/min, airway plateau pressure ≤ 30 cmH₂O).

Contraindications

There are no absolute contraindications. Relative contraindications include: advanced age; immunosuppression; cerebral hemorrhage, severe central nervous system injury, or irreversible and debilitating central nervous system pathological changes; systemic hemorrhage, contraindication to anticoagulant therapy; mechanical ventilation for more than 7 days (airway plateau pressure > 30 cmH₂O and FiO₂ > 90%).

Oxygenation management protocol for combined ventilator and venovenous extracorporeal membrane oxygenation therapy

Oxygenation management indicators during combined therapy

General guidelines

During the process of combined mechanical ventilation and VV-ECMO, oxygenation status is dynamically assessed using oxygenation management indicators. These indicators should always remain within a reasonable range of 3–5. This range is derived through laboratory measurements of the patient's blood samples, followed by calculation using Formula (1).

Calculation formula

$$\text{Oxygenation Management Index} = (\text{Hb} \times \text{SaO}_2 \times 1.34 + \text{PaO}_2 \times 0.0031) / [\text{Hb} \times (\text{SaO}_2 - \text{SvO}_2) \times 1.34] \quad (1)$$

where:

Oxygenation Management Index—The target range for oxygenation management when using ventilator and VV-ECMO should consistently be maintained within a reasonable range of 3–5.

Hb—Measured from a venous blood sample, in g/dL.

SaO₂—Measured from an arterial blood sample, in %.

PaO₂—Measured from an arterial blood sample, in mm Hg.

SvO₂—Measured from a blood sample drawn upstream of the ECMO membrane, in %.

Initial oxygenation management

Initial ventilator settings^[8]

- Initially provide a moderate to high level of PEEP (≥10 cmH₂O).
- Dynamically adjust tidal volume, respiratory rate, PEEP, and FiO₂, with consideration of such factors as alveolar

recruitment potential, risk of barotrauma, lesion homogeneity, hemodynamics, and body weight.

Initial venovenous extracorporeal membrane oxygenation settings.

1. Blood flow: 60–80 mL/kg/min. If blood volume is sufficient but venous drainage is significantly lower than this value, potential causes such as tubing, cannula placement, or intrathoracic/intra-abdominal pressure should be investigated.
2. Gas flow: Blood flow ratio $\approx 1:1$. For patients with hypercapnia and marked acidemia, gas flow may be increased appropriately.
3. Air-oxygen blender: Initial oxygen concentration is set to 100%.

Initial oxygenation indicator evaluation

After initial settings, calculate using Formula (1) to assess whether oxygenation is within the reasonable range as described in section titled “Oxygenation management indicators during combined therapy.” If so, proceed to section titled “Oxygenation management during operation”; if not, follow the steps in section titled “Oxygenation management when assessment is outside the target range.”

Oxygenation management during operation

Monitoring and management of barotrauma risk

During the operation period, evaluate for any signs of barotrauma (eg, pneumothorax, subcutaneous emphysema) or high risk of barotrauma (eg, progressive increase in peak inspiratory pressure/plateau pressure, presence of lung bullae on chest imaging). If present, implement ventilator protective ventilation settings. Simultaneously, consider adjusting and re-assessing VV-ECMO parameters.

Ventilator ultraprotective ventilation parameter settings^[9, 10]

Parameter settings are as follows:

1. FiO_2 : 30%–50%.
2. Tidal volume: Tidal volume ≤ 4 mL/kg PBW.
3. PEEP: Maintain PEEP ≥ 10 cmH₂O during VV-ECMO support. Individualize settings based on transpulmonary pressure monitoring, lung electrical impedance tomography, lung ultrasound, etc.
4. Respiratory rate: 4–15 breaths/min.
5. Plateau pressure: Keep <25 cmH₂O.
6. Driving pressure: ≤ 15 cmH₂O. May be increased slightly in pulmonary fibrosis or reduced chest wall compliance.

Venovenous extracorporeal membrane oxygenation parameter adjustment^[11, 12]

VV-ECMO parameter adjustments are as follows:

1. Hypoxemia and hypercapnia should be prevented during combined use. The gas exchange targets are PaO_2 : 65–90 mm Hg, PaCO_2 : 35–45 mm Hg.
2. Avoid rapid correction of hypercapnia after starting VV-ECMO to prevent respiratory alkalosis. PaCO_2 should not be reduced by more than 10 mm Hg per hour. After 24 hours, PaCO_2 should not decrease more than 50% of its pre-VV-ECMO level to prevent neurological complications.

Oxygenation metric assessment during operation

During operation, Formula (1) is used to assess whether oxygenation is within the reasonable range described in section

titled “Oxygenation management indicators during combined therapy.” If within range, proceed to section titled “Initial oxygenation management”; if not, follow the steps in section titled “Oxygenation management when assessment is outside the target range.”

Oxygenation management when assessment is outside the target range

VV-ECMO parameter adjustments: Prioritize adjusting VV-ECMO parameters, such as blood flow, oxygen concentration, and gas flow rate.

Ventilator parameter adjustments: If the index calculated per Formula (1) still does not fall within the reasonable range after adjusting VV-ECMO parameters, increase the ventilator FiO_2 , appropriately increase tidal volume, adjust PEEP and respiratory rate, and then recalculate and assess using Formula (1) until the index reaches the reasonable range.

Concurrently with oxygenation adjustments, actively identify the causes of inadequate oxygen supply and administer targeted treatment for these causes.

Oxygenation management during the weaning process

Once the patient’s primary condition has been controlled and with stable ventilator settings, gradually reduce VV-ECMO support. Using Formula (1), evaluate whether oxygenation reaches the reasonable range. If within range and the weaning criteria are met, proceed with a VV-ECMO weaning trial. Adjust ventilator parameters as necessary during the weaning trial. VV-ECMO weaning criteria^[13]: (1) Effective control of the primary condition and improved findings in lung imaging. (2) $\text{PaO}_2 \geq 70$ mm Hg with $\text{FiO}_2 \leq 60\%$ and $\text{PEEP} \leq 10$ cmH₂O under ventilator support. (3) With ventilator settings of tidal volume 6–8 mL/kg PBW, plateau pressure ≤ 28 cmH₂O, and respiratory rate ≤ 28 breaths/min, arterial blood gas analysis should show pH 7.35–7.45, and PaCO_2 35–45 mm Hg, with no excessive respiratory effort from the patient.

VV-ECMO weaning trial protocol: (1) Oxygen reduction: When the patient’s lung function begins to improve, the VV-ECMO oxygen concentration should be tapered from 100% to 21%, decreasing by approximately 20% each time. Assess whether SpO_2 remains $>92\%$ or PaO_2 remains ≥ 70 mm Hg after each reduction. If the criteria are met, proceed with step 2. If not, pause the weaning trial. (2) Gas flow reduction: Gradually reduce the VV-ECMO gas flow from its current setting to zero, at a rate of 0.5–1 L/min each time. After each adjustment, observe for 30 minutes and repeat arterial blood gas analysis. Assess PaO_2 , pH, and PaCO_2 to see if they meet the criteria of $\text{PaO}_2 \geq 70$ mm Hg, pH 7.35–7.45, and PaCO_2 35–45 mm Hg, and if the patient exhibits no excessive respiratory effort. If these criteria are met, proceed with step 3; if not, pause the weaning trial. (3) Clamp gas line: Maintain the gas flow cessation for a minimum of 2–3 hours, which may need to be extended to 24 hours or longer for some patients. Subsequently, repeat arterial blood gas analysis and evaluate arterial PaO_2 , pH, and PaCO_2 to assess if they meet the criteria of $\text{PaO}_2 \geq 70$ mm Hg, pH 7.35–7.45, and PaCO_2 35–45 mm Hg, and if the patient exhibits no excessive respiratory effort. If these criteria are met, proceed with step (4); if not, pause the weaning trial. (4) Decannulation: Following the steps described above, ECMO removal can proceed if the evaluation criteria are met after step (3); if not, pause the weaning trial.

Ventilator parameter adjustments^[14]: During the weaning trial, gradually adjust ventilator parameters to achieve tidal volume 6–8 mL/kg PBW, plateau pressure ≤ 28 cmH₂O, and respiratory rate ≤ 28 breaths/min. After VV-ECMO withdrawal, continue respiratory support according to standard principles.

Oxygenation regulation following failed VV-ECMO weaning trial

If the weaning trial fails, the VV-ECMO and ventilator settings are restored to the parameters used during the VV-ECMO operation period (see section titled “Oxygenation management during operation”), and dynamic reassessment should be continued. If the weaning criteria are met, the weaning trial is repeated until successful withdrawal.

This document was drafted in accordance with GB/T 1.1—2020 *Directives for standardization—Part 1: Rules for the structure and drafting of standardizing documents*.^[15]

Please note that certain content within this document may involve patent rights. The issuing organization shall not be held responsible for identifying any or all such patent rights.

Principal drafters

Chuanzhu Lv (Sichuan Academy of Medical Sciences & Sichuan Provincial People's Hospital), Wei Guo (Beijing Hospital of Traditional Chinese Medicine, Capital Medical University), Xufeng Chen (The First Affiliated Hospital of Nanjing Medical University), Yong Mei (The First Affiliated Hospital of Nanjing Medical University), Jing Sun (First Medical Center, Chinese PLA General Hospital), Ling Wang (Shenzhen Traditional Chinese Medicine Hospital), Yuguo Chen (Qilu Hospital of Shandong University), Huadong Zhu (Peking Union Medical College Hospital), Yu Cao (West China Hospital, Sichuan University), Yanfan Chai (Tianjin Medical University General Hospital), Xin Li (Guangdong Provincial People's Hospital), Changju Zhu (The First Affiliated Hospital of Zhengzhou University), Haiyan Zhu (First Medical Center, Chinese PLA General Hospital), Wei Gu (Beijing Chuiyangliu Hospital), Wensheng Qi (Guang'anmen Hospital, China Academy of Chinese Medical Sciences), Xianliang Yan (The Affiliated Hospital of Xuzhou Medical University), Peng Peng (The First Affiliated Hospital of Xinjiang Medical University), Shuyin Sun (The First People's Hospital of Jining City), Honghong Pei (The Second Affiliated Hospital of Xi'an Jiaotong University), Li Jiang (Xuanwu Hospital, Capital Medical University), Qingbian Ma (Peking University Third Hospital), Yu Ma (Chongqing University Central Hospital/Chongqing Emergency Medical Center), Yun Han (Guangdong Provincial Hospital of Traditional Chinese Medicine), Wei Zhang (The First Affiliated Hospital of Kunming Medical University), Hengbo Gao (The Second Hospital of Hebei Medical University), Zhongqiu Lu (The First Affiliated Hospital of Wenzhou Medical University), Xiaotong Han (Hunan Provincial People's Hospital), Wei Chong (The First Hospital, China Medical University), Jing Yan (Zhejiang Hospital Affiliated to Medical College of Zhejiang University), Wei Chen (Beijing Shijitan Hospital, Capital Medical University), Bei Hu (Guangdong Provincial People's Hospital), Xizhen Fan [The First Affiliated Hospital of University of Science and Technology of China (Anhui Provincial Hospital)], Deya Shang (Shandong Provincial Hospital), Chunshui Cao (The First Affiliated Hospital of Nanchang University), Yucai Hong (Sir Run Run Shaw Hospital, Zhejiang University School of Medicine), Peiwu Li (Lanzhou University Second Hospital),

Jianqi Lu (The First Affiliated Hospital of Guangxi University of Chinese Medicine), Yang Chen (The First Affiliated Hospital of Guizhou University of Traditional Chinese Medicine), Jun Duan (China-Japan Friendship Hospital), Tengfei Chen (Beijing Hospital of Traditional Chinese Medicine, Capital Medical University), Jian Xu (RocketMedical Co., Ltd.), Chuanzheng Sun (Third Xiangya Hospital, Central South University), Yuxiang Liu (Shenzhen Traditional Chinese Medicine Hospital), Xueyan Li (Beijing Hospital of Traditional Chinese Medicine, Capital Medical University), Yan Zhao (Zhongnan Hospital of Wuhan University), Mei Meng (Ninth People's Hospital, Shanghai Jiao Tong University School of Medicine), Yue Han (Shenzhen Traditional Chinese Medicine Hospital), Mujuan Xu (Shenzhen Traditional Chinese Medicine Hospital), Chenqi Zhu (RocketMedical Co., Ltd.).

Conflict of interest statement

Yuguo Chen is Editor-in-Chief of *Emergency and Critical Care Medicine*. Chuanzhu Lv and Yan Zhao are Associate Editors of *Emergency and Critical Care Medicine*. Xin Li and Qingbian Ma are Editorial Board Members of *Emergency and Critical Care Medicine*. Wei Gu is a Youth Editorial Board Member of *Emergency and Critical Care Medicine*. The article was subject to the journal's standard procedures, with peer review handled independently of the Editor-in-Chief, Associate Editors, Editorial Board Members, the Youth Editorial Board Member, and their research groups. The authors declare no conflict of interest.

Author contributions

Principal drafters listed earlier contributed to the draft of the manuscript.

Funding

None.

Ethical approval of studies and informed consent

Not applicable.

Acknowledgements

None.

References

- [1] WS 392-2012. *Clinical Application of Mechanical Ventilators*. Ministry of Health of the People's Republic of China; 2012.
- [2] National Health Commission of the People's Republic of China. Technical operation standard for adult extracorporeal membrane oxygenation (2024 edition). Accessed September 30, 2025. <https://www.nhc.gov.cn/yzygj/c100068/202411/2ae6596f7a9a4c889f1b0004a6f3b543.shtml>
- [3] Professional Committee of Critical Care Medicine, Respiratory Physicians Branch, Chinese Medical Doctor Association; Critical Care Group, Respiratory Disease Branch, Chinese Medical Association. Recommendations for extracorporeal membrane oxygenation (ECMO) in adult severe respiratory failure. *Chin J Tuberc Respir Dis*. 2019;42(9):660–684.
- [4] Anesthesia Committee of Chengdu Medical Association, Chongqing Medical Association Anesthesiology Branch. Perioperative mechanical ventilation for adult non-cardiothoracic surgical patients: a Delphi consensus. *Huaxi Med*. 2024;39(11):1979–1690. doi:10.7507/1002-0179.202410028

- [5] Brodie D, Peek G, MacLaren G, Vercaemst L, Thiagarajan R. *Extracorporeal Life Support: The ELSO Red Book, 6th Edition*. Extracorporeal Life Support Organization; 2022.
- [6] Assouline B, Combes A, Schmidt M. Setting and monitoring of mechanical ventilation during venovenous ECMO. *Crit Care*. 2023;27(1):95. doi:10.1186/s13054-023-04372-2
- [7] National Health Commission of the People's Republic of China. Notice on issuing the Operational Specifications for Adult Extracorporeal Membrane Oxygenation (2024 Edition). Accessed August 20, 2025. http://www.nhc.gov.cn/yzygj/c100068/202411/2ae6596f7a9a4c889f1b0004a6f3b543/files/1732873199940_40343.pdf
- [8] National Health Commission of the People's Republic of China. Notice of the general office of the National Health Commission on issuing the National Catalogue of Restricted Medical Technologies and the Specifications for Clinical Application (2022 edition). Accessed August 20, 2025. <https://www.nhc.gov.cn/yzygj/c100068/202204/2655831f6f444b00b3e50604e67531f5.shtml>
- [9] Tonna JE, Abrams D, Brodie D, et al. Management of adult patients supported with venovenous extracorporeal membrane oxygenation (VV-ECMO): guideline from the Extracorporeal Life Support Organization (ELSO). *ASAIO J*. 2021;67(6):601–610. doi:10.1097/MAT.0000000000001432
- [10] Combes A, Hajage D, Capellier G, et al; EOLIA Trial Group, REVA, and ECMONet. Extracorporeal membrane oxygenation for severe acute respiratory distress syndrome. *N Engl J Med*. 2018;378(21):1965–1975. doi:10.1056/NEJMoa1800385
- [11] Quintel M, Busana M, Gattinoni L. Breathing and ventilation during extracorporeal membrane oxygenation: how to find the balance between rest and load. *Am J Respir Crit Care Med*. 2019;200(8):954–956. doi:10.1164/rccm.201906-1164ED
- [12] Joyce CJ, Udy A, Weeden M, Shekar K, Cook DA. What determines the arterial partial pressure of carbon dioxide on venovenous extracorporeal membrane oxygenation? *ASAIO J*. 2022;68(8):1093–1103. doi:10.1097/MAT.0000000000001604
- [13] Cavayas YA, Munshi L, Del Sorbo L, Fan E. The early change in PaCO₂ after extracorporeal membrane oxygenation initiation is associated with neurological complications. *Am J Respir Crit Care Med*. 2020;201(12):1525–1535. doi:10.1164/rccm.202001-0023OC
- [14] Abrams D, Schmidt M, Pham T, et al. Mechanical ventilation for acute respiratory distress syndrome during extracorporeal life support. Research and practice. *Am J Respir Crit Care Med*. 2020;201(5):514–525. doi:10.1164/rccm.201907-1283CI
- [15] State Administration for Market Regulation, National Standardization Administration. Directives for standardization—Part 1: rules for the structure and drafting of standardizing documents. Accessed September 30, 2025. <https://openstd.samr.gov.cn/bzgk/gb/newGbInfo?hcno=C4BFD981E993C417EF475F2A19B681F1>

How to cite this article: Lv C, Guo W, Chen X, et al. Oxygenation adjustment of ventilator combined with venovenous extracorporeal membrane oxygenation. *Emerg Crit Care Med*. 2026;6(1):65–69. doi: 10.1097/EC9.0000000000000174